




Why operator messages are not shown anymore?

Microplan software Parseq is designed to manage multiple languages as regard as the messages shown to the operator during the execution of test sequences.

The message "This page can't be displayed" might be shown on screen if customer changes the language and there is no translation associated to the selected language.



To solve this problem you should change the language with the drop-down menu in the bottom right hand corner of the screen, as shown in the following picture.

For instance change from Italian to the language that was in use prior to the problem, the one used to display messages to the operator.

Language change takes effect with the following operator message so, after selecting the new language, you should proceed with the sequence, in order to let a new operator message be displayed on the screen.

